

TEMP (°C)	RESIS (Ω)	TEMP (°C)	RESIS (Ω)	TEMP (°C)	RESIS (Ω)	TEMP (°C)	RESIS (Ω)	TEMP (°C)	RESIS (Ω)	TEMP (°C)	RESIS (Ω)	TEMP (°C)	RESIS (Ω)
-55	9,637,300	-22	1,091,100	11	189,700	44	45,430	77	13,850	104	6,046	130	3,010
-54	8,952,300	-21	1,029,000	12	180,890	45	43,680	78	13,400	105	5,876	131	2,936
-53	8,320,400	-20	970,720	13	172,540	46	42,010	79	12,970	106	5,712	132	2,863
-52	7,737,100	-19	916,130	14	164,620	47	40,420	80	12,550	107	5,553	133	2,793
-51	7,198,600	-18	864,940	15	157,110	48	38,890	81	12,150	108	5,399	134	2,724
-50	6,700,900	-17	816,890	16	149,980	49	37,420	82	11,770	109	5,250	135	2,658
-49	6,240,900	-16	771,830	17	143,220	50	36,020	83	11,400	110	5,106	136	2,594
-48	5,815,400	-15	729,510	18	136,800	51	34,680	84	11,040	111	4,967	137	2,531
-47	5,421,700	-14	689,770	19	130,710	52	33,400	85	10,700	112	4,832	138	2,471
-46	5,057,100	-13	652,420	20	124,920	53	32,170	86	10,370	113	4,702	139	2,411
-45	4,719,400	-12	617,320	21	119,420	54	30,990	87	10,050	114	4,575	140	2,354
-44	4,406,300	-11	584,310	22	114,190	55	29,860	88	9,738	115	4,452	141	2,298
-43	4,116,000	-10	553,260	23	109,220	56	28,780	89	9,441	116	4,334	142	2,244
-42	3,846,700	-9	524,040	24	104,500	57	27,740	90	9,154	117	4,219	143	2,191
-41	3,596,700	-8	496,540	25	100,000	58	26,750	91	8,878	118	4,108	144	2,140
-40	3,364,500	-7	470,640	26	95,720	59	25,790	92	8,611	119	3,999	145	2,090
-39	3,148,800	-6	446,250	27	91,650	60	24,880	93	8,354	120	3,896	146	2,042
-38	2,948,200	-5	423,260	28	87,770	61	24,000	94	8,106	121	3,794	147	1,994
-37	2,761,700	-4	401,580	29	84,080	62	23,160	95	7,866	122	3,696	148	1,949
-36	2,588,200	-3	381,150	30	80,570	63	22,350	96	7,635	123	3,601	149	1,904
-35	2,426,600	-2	361,870	31	77,220	64	21,570	97	7,412	124	3,509	150	1,861
-34	2,276,100	-1	343,680	32	74,030	65	20,830	98	7,196	125	3,419		
-33	2,135,900	0	326,500	33	70,980	66	20,110	99	6,987	126	3,332		
-32	2,005,200	1	310,290	34	68,080	67	19,410	100	6,786	127	3,248		
-31	1,883,300	2	294,980	35	65,310	68	18,760	101	6,591	128	3,167		
-30	1,769,600	3	280,510	36	62,680	69	18,130	102	6,403	129	3,087		
-29	1,663,400	4	266,840	37	60,160	70	17,520	103	6,221				
-28	1,564,300	5	253,910	38	57,750	71	16,930						
-27	1,471,600	6	241,690	39	55,460	72	16,360						
-26	1,385,000	7	230,120	40	53,260	73	15,820						
-25	1,304,100	8	219,170	41	51,170	74	15,300						
-24	1,228,400	9	208,800	42	49,170	75	14,790						
-23	1,157,500	10	198,990	43	47,260	76	14,310						

---	ISO RELEASE	04/24/03	DD
REV	REV RECORD	DATE	APP

SCALE	NONE
BY	DAN DANKERT
DATE	10/26/99
REV.	NONE
LAYER	0 OF 1

U.S. SENSOR CORP.
1832 W. COLLINS AVE.
ORANGE, CA. 92867

RESISTANCE-TEMPERATURE TABLE

100,000 Ω "J"